

(TO BE SUBMITTED ALONG WITH TECHNICAL BID)

Page 1

TENDER NOTIFICATION NO: -----

Phone No:- 27296326

TENDER FORM

The Registrar,
Delhi Technological University,
Bawana Road,
Delhi-110042

We the undersigned (herein after called as Contractor/Vendors/Suppliers) hereby offer to execute supply of items as per specification against which we have quoted over rates and for which this tender may be accepted at the rates stated there in and subject to the terms & conditions set forth for such items as may be ordered by the Registrar, Delhi Technological University or officer acting on his behalf.

Date this _____ Day of _____

Signature of Contractor _____

Address _____

INDEX

S.No.	Particulars of documents	No. of pages

Guidelines/Procedure to be followed in introduction of 'e'-procurement solution:

1. Payment of cost of Tender documents: The collection of cost of Tender documents is dispensed away with, as there is no physical supply of tender documents and also to have absolute anonymity of bidder participating in e-procurement solution. The bidders can view/download the tender documents from the: <https://govtprocurement.delhi.gov.in> .

2. Submission of bids: The bidders who are desirous of participating in 'e'-procurement shall submit their price bids in the standard formats prescribed in the Tender documents, displayed at : <https://govtprocurement.delhi.gov.in>. The bidder should upload the scanned copies of all the relevant certificates, documents etc. in the: <https://govtprocurement.delhi.gov.in>. in support of their price bids. The bidder shall sign on all the statements, documents, certificates, uploaded by him, owning responsibility for their correctness/authenticity and copies thereof may also be submitted in the office of the Asst. Registrar(S&P), DTU along with original EMD. However, documents of the bidders downloaded online or requisitioned subsequently only will form the basis for deciding the tender.

3. Payment of Bid Security (Earnest Money Deposit): The EMD shall be in the form of the Demand Draft/Pay Order/BG/Fixed Deposit Receipt of a scheduled bank issued in favour of Registrar, Delhi Technological University, Delhi and the zerox copy thereof is to be scanned and uploaded along with the bid, and the original shall be sent to DTU so as to reach before the date & time of closing of the bids; failing which bid will be rejected.

4. Price Bid Opening: The Price Bids will be opened online by the concerned officer/officers at the specified date & time and the result will be displayed on the: <https://govtprocurement.delhi.gov.in>. which can be seen by all the bidders who participated in the tenders.

5. Processing of Tenders: The concerned officer/officers will evaluate and process the tenders as done in the conventional tenders and will communicate the decision to the bidder online.

6. Payment of Performance Guarantee: The successful tenderer shall furnish a bank guarantee/FDR of the value of 05% of the basic cost of the item for a period of 60 days beyond the warranty period from a nationalized bank to ensure the satisfactory performance of item supplied. The performance guarantee is to be submitted at the time of installation / demonstration of equipments. In case the performance of the item is not found satisfactory, the amount of bank guarantee will be forfeited & credited in university account.

7. Participation of Bidders at the time of opening of bids: Bidders have two options to participate in tendering process at the time of opening of Bids:

- (i). Bidders can come at the place of opening of bids (electronically) as done in the conventional tender process.
- (ii). Bidders can visualize the process online.

8. Participation Financial Rules for e-procurement: The e-procurement system would be applicable for purchase of goods, outsourcing of services and execution of work as prescribed in GFRs.

ASSTT. REGISTRAR (S&P)
DELHI TECHNOLOGICAL UNIVERSITY,
SHAHBAD DAULATPUR, BAWANA ROAD,
DELHI – 110 042

TERMS AND CONDITIONS

TENDER FORMS ARE NOT TRANSFERABLE

Procedure for submission of bids: The bidders who are desirous of participating in 'e'-procurement shall submit their technical and price bids in the standard formats prescribed in the Tender documents, displayed at: <https://govtprocurement.delhi.gov.in>. The bidder should upload the scanned copies of all the relevant certificates, documents etc. after page-numbering all documents and tender document and prepare an index thereof in the: <https://govtprocurement.delhi.gov.in>. in support of their price bids. The bidder shall sign on all the statements, documents, certificates, uploaded by him, owning responsibility for their correctness/authenticity and copies thereof may also be submitted in the office of the Asstt. Registrar (S&P), DTU along with original EMD.

1. The Competent Authority of the University does not bind himself to accept the lowest or any tender.

2. ALTERATION IN THE SPECIFICATION.

(i) The specifications mentioned/issued with this form of tender must not be altered by the Suppliers.

3. INCOMPLETE TENDERS

Tender will not be considered if complete information is not given at the time of tendering or if the particulars and data (if any) asked for are not given.

4. CANCELLATION OF TENDER/ CONTRACT/ IN PART OR IN FULL IN CASE OF DEFAULT IN CONTRACT/SUPPLY:

If the Supplier, in the opinion of the Institute fails or neglects to comply with any of the terms & conditions forming, part of the order issued, the head of institute shall without prejudice to any other right or remedies under the contract, has the right to cancel the contract /order by giving 15 days notice in writing to the Suppliers/firms without being liable to pay compensation for such cancellation.

5. Tender shall be uploaded as per guidelines indicated for e-procurement solution.

6. Demonstration of equipments has to be arranged by the suppliers, if desired by the institute. The technical committee may visit production facility if so desired for sample verification.

7. The quotation should be valid for a period of one year from the date of opening of the tender.

8. Rates are to be quoted in INR (Rupee terms) only and any revision thereof is not allowed after the tenders have been opened.

9. The delivery period should be clearly mentioned against each item, incase, the items are not readily available, ex-stock offer will be preferred.

10. Rates should be quoted F.O.R Institution. Sales tax/VAT/Octroi, Custom duty and other taxes leviable, should be mentioned clearly; indicating whether these are to be charged extra or included in the quoted price.

11. Consignment will not be insured at the Institute / University Cost.

12. Preference will be given to quotation pertaining to indigenous products, However, where suitable substitutes are not available and item need to be imported the following clarification / information should be given.

(i) Whether the item will be imported by the intended tenderers against its own import license or university will have to provide Custom Exemption Certificate (CDE).

(ii) Name and address of the foreign supplier.

(iii) Break up of CIF, and duty (if paid) should be given along with service charges if any.

- (iv) Delivery period including information about mode of dispatch and possible duration (after dispatch) for receipt of item at the port.
 - (v) Whether the item required any special preparation for installation. In case yes, full details should be given regarding operation maintenance of the items.
 - (vi) In case of costly/sophisticated items whether the tenderers will arrange any special training regarding operation / maintenance of the items.
 - (vii) Nature of assurance for the supply of spares after the warranty period.
13. The payment will be made within 30 days after the successful demonstration/installation of the equipment. Rejected items/goods should also be removed within 30 days after which no responsibility will be accepted by University.
 14. Conditional quotations and/or incomplete quotations in any respect will be rejected.
 15. In case you cannot quote for one or more of the items asked for in the tender the word "NOT QUOTED" (in the rate column) should be indicated.
 16. The specification of the item quoted by the firm should confirm to the University specifications. Confirmation, in this respect should be specifically mentioned in the tender. Where the tenderer feels that the specification of the item not fully given or differ, from the specification of the item mentioned by the university, the exact specification of such item should be attached with the tender indicating the item quoted.
 17. The Firm is required to link the University specifications with catalogues & leaflets/literature for each item. Details features, for compliance of specification should be provided on specification sheet & appropriate reference i.e. page no. & para of literature, leaflet wherefrom the relevant information has been checked, should be indicated.
 18. **EARNEST MONEY:-** EMD should be attached with the technical bid. The EMD shall be in the form of the Demand Draft/Pay order/BG/Fixed Deposit Receipt of a scheduled bank issued in favour of Registrar, Delhi Technological University, Delhi. Zerox copy thereof is to be scanned and uploaded along with the bid, and the original instruments shall be sent to DTU so as to reach before the date of closing of the bids. Failure to furnish the original instrument before the closing of the bid, will entail rejection of bid and blacklisting. If the tenderer after acceptance of the tender refused to take up the purchase order, his Earnest Money will be forfeited. Any tender received without / less Earnest Money deposit shall be summarily rejected.
 19. The Competent Authority reserves the right to reject any or all the tenders without assigning any reason, at any stage, and his decision will be final.
 20. The supplies shall have to be made within the period specified in the purchase order failing which the order shall be cancelled and the Earnest Money will be forfeited. However, in exceptional circumstance and, on written request, from the supplier/tenderer, extension of date for supply of the material will be considered. Extension in supply period is at the sole discretion of the competent authority.
 21. Service manuals, wherever available/ required, should be provided along-with the Equipments.
 22. A **WARRANTY** certificate should invariably be supplied along with the item at the time of delivery. Non-Compliance of the same will result in non-acceptance of the item from the firm with whom the order was placed beside rejection of the tender.
 23. The Competent Authority reserves the right to levy liquidated damages up to 2% of the value of the order for delayed supply. If the supply is delayed beyond the extended period, the University reserves the right even to cancel the order and forfeit the EMD of the firm/ tenderer.
 24. **PERFORMACE SECURITY DEPOSIT:-** The successful tenderer shall furnish Performance Security Deposit of the value of 05% of the basic cost of the item in

the shape of Bank Guarantee/FDR etc. from a nationalized bank pledged to Registrar, DTU, for a period of 60 days beyond the warrantee period to ensure the satisfactory performance of item supplied. The performance guarantee is to be submitted at the time of installation / demonstration of equipments. In case the performance of the item is not found satisfactory, the amount of bank guarantee will be credited in university account.

25. DEFAULT: - In the event of default and unsatisfactory service of the contractor/Supplier firm, the DTU will be at liberty to repair/get the item serviced from other party at the cost of supplier/ contractor/ tenderer.
26. In case of software items, the suppliers should ensure that:-
 - i. Legal software is supplied in original sealed pouches / P. K. T.
 - ii. A license agreement is enclosed with it.
 - iii. A registration card is available for software.
27. FAILURE AND TERMINATION: - If the Contractor / Supplier fails to deliver the stores or any installment thereof within the period fixed for such delivery or at any time repudiates the contract before the expiry of such period, DTU may without prejudice to the right of the purchaser may recover damages for breach of the contract.
28. The technical & financial bids of only those bidders will be opened who fulfill the eligibility criteria and whose documents are found in order. If any of the date earmarked for opening of technical or financial bids happens to be holiday, the bids will be opened on the very next working day.
29. For any query/clarification in r/o technical aspect of the enquiry, HOD (Mech. Engg.), DTU may be contacted.
30. Disputes, if any, arising out of this tender shall be subject to exclusive jurisdiction of Courts of Delhi/New Delhi only.

ASSTT. REGISTRAR (S&P)
DELHI TECHNOLOGICAL UNIVERSITY,
SHAHBAD DAULATPUR, BAWANA ROAD,
DELHI – 110 042

CHECK LIST OF DOCUMENTS TO BE SENT WITH TECHNICAL BID.

S.No.	Particulars of documents	Page no.	No. of pages
1.	Proof of EMD (mention amount with instrument number and date)		
2.	Proof of PAN no. (mention no.....)		
3.	Proof VAT/Sales Tax Registration. (mention no.)		
4.	Brochure/Leaflets/Technical Information of the item(s), if any		
5.	Page no. 1 of Tender Document, duly signed.		
6.	Detailed Technical specifications, Terms & Conditions and Delivery period to be submitted on firm's letter head		
7.	Warranty Certificate		

Note: All copies of above documents should be duly signed and stamped by the tenderer before uploading.

Signature of tenderer:
Name:.....
Name of firm:
Telephone No.....

GOVERNMENT OF NATIONAL CAPITAL TERRITORY OF DELHI
 DELHI TECHNOLOGICAL UNIVERSITY
 (FORMERLY DELHI COLLEGE OF ENGINEERING) Ph. 27296326
 SHAHBAD DAULATPUR: BAWANA ROAD: DELHI-110 042

No.F.211/1-2/11-12/P/

Dated:

NOTICE INVITING TENDER

E- Tenders along with illustrated literature/leaflets for the supply/execution of item(s)/stores/work detailed below are invited in two-bids system through 'e' procurement solution only as per the guidelines and terms & conditions given in tender document - details of the NIT along with terms & conditions, specifications etc. can be seen/downloaded at/from the website.

The interested tenderers should upload duly signed tender form and their bids along with scanned copies of all the relevant certificates, documents etc. in support of their technical & price bids on the: <https://govtprocurement.delhi.gov.in>. latest by **30/01/2012 up to 3.00P.M.** **An index prepared after pagination of all documents may also be uploaded and** the technical bids will be opened online on the **same date at 3.30 P.M** for only those bidders whose draft for EMD amount is dropped in Tender Box placed in the office of Asst. Registrar (S&P) in the presence of the bidders who wish to be present and also displayed on the website. For participation in the tender through e-procurement solution, the tenderers are required to have digital certificate and get registered with application Service Provider NIC.

Tender document is also available for viewing on the website of Delhi Technological University, Delhi at www.dce.edu and www.dce.ac.in

Yours faithfully,

EMD: Rs.48,000/-

ASST. REGISTRAR (S&P),

S.No.	Item Description	Qty. (no.)
1	UTM (Table Top)50 KN Dual Column for tensile/compressive/double shear testing of materials as per ASTM, BS, DIN, ISO, EN Stndrs. For circular and flat specimens compatible with attachments, with complete Data Acquisition System Specifications on next page.....	01

Contd.....

TECHNICAL SPECIFICATIONS OF 50kN UTM- TABLE TOP

Electro-mechanically operated Dual-Column Materials Testing (Tensile and Compressive) System with digital closed loop position, load and strain controls with minimum capacity of 50 kN having the following Minimum or Better features.:

- Tension/Compression Load Cell of capacity 50 kN with a least count of 1 gms and accuracy of $\pm 0.5\%$ up to $1/100^{\text{th}}$ of load cell capacity
- Maximum width and depth should be approx. 750mm and 700mm respectively
- Total Vertical test space 1100 mm and cross head travel 1050 mm with a digital crosshead drive system having variable speed controls and crosshead speed accuracy of $\pm 0.2\%$ or lower of set speed (maximum crosshead return speed: 500 mm/min.)
- Space between the columns should be min. 400mm.
- Testing speed range: 0.05 to 500 mm/min.
- Maximum force at highest speed: 25 kN
- Maximum speed at the highest load: 250 mm/min.
- Position control resolution maximum upto 0.133 mm
- Load weighing system should meet or surpass ASTM E4, BS 1610, DIN 51221, ISO 7500/1, EN 10002-2 and AFNOR A03-501.
- Robust guidance columns with sufficient stiffness that can ensure accurate cross-head alignment and its linear motion with minimum backlash (e.g. supported on preloaded ball screw or heavy duty bearings, etc.).
- Simultaneous data sampling of load, displacement and strain data for 40 kHz and data acquisition with a minimum sampling frequency of 100 Hz at PC with Ethernet based PC transmission. Compatible to Laptop connection with data acquisition
- Digital close loop control at a minimum sampling frequency of 100 Hz
- Basic transducer conditioner cards (with > 16 bit resolutions) for load and position control-provisions of extra transducer preferred.
- Complete software control and analysis of data for all kinds of tensile/compressive/flexural tests through properly integrated interface and analysis software based on MS-WINDOWS XP latest and standard PC P-IV or higher configuration preferred
- Software should be able to generate prompted and unprompted methods for testing
- All Software, manual etc. should be in English and SI units or possible to follow. Analysis part of software should be able to produce automatic graphs and tables of analyzed load, displacement and strain data and possibility to directly export such data in Standard ASCII or CSV formats to other graphic packages such as EXCEL etc.
- Set of compression anvils with 150mm diameter of minimum 50kN capacity
- Wedge action grips of capacity 50kN should be able to withstand a temperature range of -70 to 250 degree centigrade
- Complete set of extensometer with 100% elongation and 50mm gauge length with automatic recognition of transducer by software
- Provision of at least two number of programmable keys in main display shoes functionality could be changed as per requirement
- Company providing system should calibrate the equipment to the ASTM/ISO standards.
- Precision concentricity rings and alignment pins for load string alignment should be provided.